

Fluid pressure sensor with luer-lock connections for medical applications

The **flowplus¹⁶** is a high precision fluid pressure monitor device (up to 16 bar) designed from concept to exactly fit medical device criteria by including the standardised Luer-Lock connector system and internally, a void free flow channel coated with a virtually inert FDA compliant elastomer.

In particular, the universal single size Luer-Lock system enables equipment manufacturers to combine almost all related devices with each other. This means that the **flowplus¹⁶** may be instantly connected to syringes, drain tubes, infusion tubes, catheters, injection needle and similar, enabling any pressure fluctuations to be identified at an early stage, e.g. needle clogging, micro-bubbles in the fluid or variable pressure feed. The ISO 594 standardised Luer-Lock system features a 6° cone with 180° twist lock and has been selected by medical bodies throughout the world to facilitate rapid and reliable equipment interconnectivity.

Safety is thus greatly enhanced since the **flowplus¹⁶** can be quickly fitted with confidence, to provide feedback into monitoring and/or control systems, while operating with flows of up to 100ml/min at pressures of 0-16 bar and between 15°C to 45°C.

Consequently, when medical processes must be controlled or monitored, e.g. in the continuous conveying of liquids, then the **flowplus¹⁶** is easy to fit and being pre-calibrated it is exceptionally easy to use. Pressure evaluation up to 16 bar is possible with a high degree of precision via a standardised 0.1 to 10 VDC output signal. This makes the **flowplus¹⁶** a simple “fit-and-rely-on” item for monitoring process safety parameters.

Fluid pressure sensor with luer-lock connections for medical applications



Supplied by:

intertronics

INTERTRONICS

12a Station Field Industrial Estate, Banbury Road, Kidlington
Oxfordshire England OX5 1JD

t 01865 842842 e info@intertronics.co.uk

Fluid pressure sensor with luer-lock connections for medical applications

Last updated: November 2017

Statements, technical information and recommendations contained herein are based on tests we believe to be reliable but they are not to be construed in any manner as warranties expressed or implied. The user shall determine the suitability of the product for his intended use and the user assumes all risk and liability whatsoever in connection therewith.